



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMR-AM-pw-e-CA-LW-05-PN*
 S.N.: *4615*
 Date: *DEC 6 88*



Point $R_L = 50\Omega$
 $PRF = 5 \text{ kHz}$



Point $R_L = 50\Omega$
 $PRF = 100 \text{ kHz}$

- a) Output signal amplitude:
0 TO $\pm 205 \text{ VOLTS}$
- b) Pulse width:
50 NSER TO
- c) Rise time: *1.0 NSER*
 $\leq 10 \text{ NSER}$
- d) Fall time:
 $\leq 10 \text{ NSER}$
- e) PRF:
0 TO 100 kHz
(0.5% MAX. DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 Hz
[Signature]