


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI - CI*
S.N.: *4622*
Date: *DEC 21 88*



- a) Output signal amplitude:
≥ +40 VOUT
- b) Pulse width:
1 TO 100 NS
- c) Rise time: (SEE FILE 192)
≤ 300 PSEC
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 1000 Hz
(SEE FILE 192)
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 V OUT
150 mA (max)

50 dB ATTEN
32 VOUT/DIV
50 kHz