


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVI-61*
S.N.: *4624*
Date: *DEC 21 88*



- a) Output signal amplitude:
7.40 VOLTS
- b) Pulse width:
1 TO 100 NS
- c) Rise time: *(SEE FIG 1A)*
≤ 300 PSEK
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
(SEE FIG 1A)
OK
- g) Prime power:
±15 ±1.0 VOLT
150 mA (max)

50 dB ATTEN
32 VOLTS/DIV
50 kHz



[Signature]