


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-L1*  
S.N.: *4625*  
Date: *DEC 21 88*



- ) Output signal amplitude:  
*± 40 VOLTS*
- ) Pulse width:  
*1 TO 100 NS*  
*(SEE FIG 1A)*
- ) Rise time:  
*≤ 300 pS*
- ) Fall time:  
*≤ 2 NS*
- ) PRF:  
*0 TO 10 MHz*  
*(SEE FIG 1A)*
- ) Jitter, stability:  
*OK*
- ) Prime power:  
*OK*

*50 dB ATTEN*  
*32 VOLTS/DIV*  
*50 KHz*

*+15 ± 1.0 VOLTS*  
*150 mA (max)*