



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM14-2-C-7N-AM-1M*
 S.N.: *4630*
 Date: *NOV 30 88*



Ⓐ *40 dB ATTN*
≈ 10 VOLTS/DIV
= 1 MHz



Ⓑ *AS Ⓐ BUT*
≈ 25 MHz

- a) Output signal amplitude:
0 TO ± 300 VOLTS
- b) Pulse width:
≤ 400 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 Hz