


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV4-MPT-TM-P
S.N.: 4631
Date: DEC 15 88



50 db ATTEN
≈ 32 VOLTS/DIV
500 kHz

- a) Output signal amplitude: $> 130 \text{ VOLTS}$
- b) Pulse width: $\leq 1 \text{ nSEC}$
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1 MHz
- f) Jitter, stability: OK
- g) Prime power: $13 \pm 1.0 \text{ VOLT}$
 150 mA (max)