


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-1-N-M*
S.N.: *4659*
Date: *DEC 6 '88*



- a) Output signal amplitude:
0 TO -250.0
- b) Pulse width:
≤ 4 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN

100 VOLTS/DIV
20 KHZ PRF.

±15 ±1.0 V

250 mA

(MAX)
[Signature]