


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMR-A-1-PW-C-EXT-EN-PN*

S.N.: *4662*

Date: *DEC 16 88*



*Ⓐ P out*  
*1.0 KHz*  
*R<sub>L</sub> = 50Ω*



*Ⓑ N out*  
*1.0 KHz*  
*R<sub>L</sub> = 50Ω*

- a) Output signal amplitude:  
*0 to ± 200 V. O. A. B*
- b) Pulse width:  
*50 nsec to 1.0 μsec*
- c) Rise time:  
*6.5% DUTY CYCLE (MAX)*  
*≤ 10 nsec*
- d) Fall time:  
*≤ 10 nsec*
- e) PRF:  
*0 to 100 KHz*  
*(0.5% DUTY CYCLE (MAX))*
- f) Jitter, stability:  
*(MAX)*
- g) Prime power:  
*OK*

*120 / 240 V*  
*50-60 Hz*