


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-HV1-MA2

S.N.: 4665

Date: JAN 12 89



60 db atten
100 VOLT/DIV
20 KHZ (PKA)

- a) Output signal amplitude:
7 + 100 VOLTS
- b) Pulse width:
650 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLTS
150 mA (MAX)