


PULSE GENERATOR
PERFORMANCE CHECK

Model: AN-5-121-P
S.N.: 4666
Date: JAN 12 89



40 dB ATTEN
10 VOLTS/DIV
500 KHZ
(PRF)

- a) Output signal amplitude:
0 TO +40 V
 - b) Pulse width:
≤ 350 PSEC (FWHM)
 - c) Rise time:
NA
 - d) Fall time:
NA
 - e) PRF:
0 TO 1.0 MHz
 - f) Jitter, stability:
OK
 - g) Prime power:
+15 ± 1.0 V
150 mA (max)
- AA