


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-50-CP*

S.N.: *4670*

Date: *DEC 12 1988*



$R_L = 3 \Omega$
 $\approx 6 \text{ AMPS/DIV}$
 $PRF = 300 \text{ Hz}$

- a) Output signal amplitude:
0 TO +30 AMP TO
- b) Pulse width: *3 μ S*
0.1 TO 5.0 μ SEC
- c) Rise time:
 $\leq 40 \text{ NSEC}$
- d) Fall time:
 $\leq 40 \text{ NSEC}$
- e) PRF:
0 TO 300 Hz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 Hz.

[Signature]