


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PRR-A-2-pw-c-05-N-IM*

S.N.: *4688*

Date: *JAN 13 '89*



*50 db ATTEN
0.32 VOLTS DIV
PRF = 1.0 MHz*



*③ R_L = 50 Ω
PRF = 100 kHz*

- a) Output signal amplitude:
0 TO - 50 VOLTS
- b) Pulse width:
20 TO 200 NS
- c) Rise time:
*(20% MAX DUTY CYCLE)
≤ 5 NS*
- d) Fall time:
≤ 5 NS
- e) PRF:
*0 TO 10 MHz
(20% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power:

*120/240 V
50-60 Hz*

[Signature]