



PULSE GENERATOR
PERFORMANCE CHECK

Model: *MMP-2-C-NPN-M-OT*
 S.N.: *4698*
 Date: *APR 28 89*



Ⓐ *40 db ATTEN*
10 VOLTS/DIV
PRF = 10 MHz



Ⓑ *AS Ⓐ BUT*
1.0 nSEC/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO 10 VOLTS
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]