



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANPP-2-C-P -JM*  
S.N.: *4714*  
Date: *FEB 27 89*



A RANGE 40 dB ATTEN  
10 VOLT/DIV  
PRF = 100 KHz



B RANGE 40 dB ATTEN  
10 VOLT/DIV  
PRF = 100 KHz

- a) Output signal amplitude:  
*0 TO +20 VOLTS*
- b) Pulse width:  
*0.4 TO 100 NS*
- c) Rise time:  
*≤ 200 PSEC*
- d) Fall time:  
*≤ 300 PSEC*
- e) PRF:  
*0 TO 100 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50 - 60 Hz*

*[Signature]*