


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVA-5-1-C-P*

S.N.: *4715.*

Date: *JAN 17' 99*



40 db Attenu
10 VOLTS/DIV
100 KHz PRF

- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
≤ 130 PSEZ.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 Hz.

[Signature]