


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-1-PW-PS-N-CA*

S.N.: *4725*

Date: *FEB 22 89*



Z_o = 50 Ω
PRF = 500 MHz

- a) Output signal amplitude:
0 TO -200 V
- b) Pulse width:
0.1 TO 5.0 μS
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
0 TO 1.0 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 MHz