


PULSE GENERATOR  
PERFORMANCE CHECK

Model: PVL-2-PS-M-N

S.N.: 4740

Date: JAN 24 1989



70 db ATTEN

≈ 320 VOLTS/DIV

PRF = 1.0 KHz

- a) Output signal amplitude:  
0 TO -350 VOLTS
- b) Pulse width:  
5 TO 100 NS
- c) Rise time:  
≤ 2 NS
- d) Fall time:  
≤ 3 NS
- e) PRF:  
0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:

120/240V

50.60 Hz

[Signature]