


PULSE GENERATOR
PERFORMANCE CHECK


Model: *MI-V-C-P*

S.N.: *4745*

Date: *FEB 28 89*



Ⓐ *56 db ATTEN*
232 VOLTS/DIV



Ⓑ *AS Ⓐ BUT*
1 NS/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width:
2 TO 100 NS
- c) Rise time:
≤ 350 PSEC
- d) Fall time:
52 NS
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
170/240V
50-60 Hz

[Signature]