


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-S-C-P

S.N.: 4754

Date: MAR 14 '89



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
≤ 500 PSECS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz

40 db ATTEN
10 VOLTS/DIV
100 KHz
(PRF)

- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]