


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AV6-3-N-DU*

S.N.: *4757*

Date: *JAN 26 09*



Ⓐ *60 db ATTEN*
100 VOLTS/DIV
PRF = 5.0 KHz



Ⓑ *AS. Ⓐ BUT*
50 NS/DIV.

- a) Output signal amplitude: *> -500 VOLTS*
- b) Pulse width: *TO 50 ns*
≤ 5 nsec
- c) Rise time: *CHAMP power*
- d) Fall time: *NA*
- e) PRF: *0 TO 20 KHz.*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 ± 1.0 VOLT.*
300 mA (MAX).

