


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-5A-C-P*

S.N.: *4760*

Date: *APRIL 26 89*



*R<sub>o</sub> = 0.7  $\Omega$*

*I = 14.3 Amps /  $\mu$ DIV*

*PRF = 5.0 kHz*

- a) Output signal amplitude:  
*0 TO +40 AMPS TO*
  - b) Pulse width: *0.75  $\mu$ s*
  - c) Rise time: *20 TO 150 NSER*
  - d) Fall time:  *$\leq$  5 NSER*
  - e) PRF:  *$\leq$  5 NSER*
  - f) Jitter, stability: *0 TO 5 KHz*
  - g) Prime power: *OK*
- 120/240 V*  
*50-60 Hz*  
*[Signature]*