


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-5-C-P-05*

S.N.: *4762*

Date: *FEB 23 89*



40dB ATTEN
10 VOLT/DIV
500 kHz PAST

- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
≤ 500 psec
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]