


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVPP-2-PS-17-CA-AN*

S.N.: *4774*

Date: *APRIL 10 89*



*A out 40 db ATTEN
10 VOLTS/DIV
50 KHZ PRF*



*B out 40 db ATTEN
10 VOLTS/DIV
50 KHZ PRF*

- a) Output signal amplitude:
0 TO + 20 VOLTS
- b) Pulse width:
0.4 TO 100 NS
(TWO RANGES)
- c) Rise time:
≤ 200 PSEC
- d) Fall time:
≤ 300 PSEC
- e) PRF:
0 TO 100 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 HZ