


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-P*
S.N.: *4783*
Date: *MAR 31 89*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 6.0 nSec
- c) Rise time:
≤ 100 pSec
- d) Fall time:
≤ 135 pSec
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+24 ± 1.0 V
300 mA (max)

40 dB ATTEN
5 VOLTS/DIV
PRF = 7.0 MHz