


PULSE GENERATOR
PERFORMANCE CHECK

Model: 1720-5B-PS-P

S.N.: 4791

Date: Feb 15 1989



a) Output signal amplitude:

0 to +4% Amp

b) Pulse width:

25 to 500 nsec

c) Rise time:

≤ 10 nsec

d) Fall time:

≤ 10 nsec

e) PRF:

0 to 1.0 KHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz

$R_L = 0.7 \Omega$

10 00V/3 DIV

700 Hz (PRF)