


PULSE GENERATOR

PERFORMANCE CHECK


Model: *ANL-B3-C-P-4A*

S.N.: *4807*

Date: *FEB 10 '89*



① *50db ATTEN*
~ 32 VOLTS/DIV
PRF = 20 KHZ



③ *AS ① BUT*
V_{OUT} ≈ 30 VOLTS

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width: *70 50 μs*
- c) Rise time: *20 TO 200 NS*
≤ 5 NS
- d) Fall time: *≤ 5 NS*
- e) PRF: *0 TO 20 KHZ*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 KHZ
- h) PROP DELAY
30 ± 1.0 NS
- i) GAIN
V_{OUT} = 100 ± 1.0 VOLTS
PRV V_{EA} = +9.7 VOLTS