


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-4B-CPN*


S.N.: *4819*

Date: *JAN 22 1996*



- a) Output signal amplitude:  
*0 TO ± 400 VOLTS*
- b) Pulse width: *10 > 50 ns*  
*0.1 TO 100 us*
- c) Rise time: *0.5% MAX DUTY*  
*≤ 20 ns*
- d) Fall time: *≤ 20 ns*
- e) PRF: *0 TO 1 kHz*  
*0.5% MAX DUTY*  
*CYCLE*
- f) Jitter, stability:  
*OK*
- g) Prime power: *170/240 V*  
*50-60 Hz*

ⓐ *P<sub>out</sub>, R<sub>L</sub> = 50 Ω*



ⓑ *N<sub>OUT</sub> R<sub>L</sub> = 50 Ω*

