


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-3B-P-C-PUI*

S.N.: *4820*

Date: *MAY 18 89*



*40 db ATTEN*

*10 VOLTS/DIV*

*R = 2 Ω*

*≈ 5 AMPS/DIV*

*PRF = 20 kHz*

- a) Output signal amplitude: *+1 TO +10 AMPS TO*
- b) Pulse width: *≈ 2 μs*  
*≤ 700 psec AT HALF*
- c) Rise time: *≤ 1.2 nsec AT 10% AND*
- d) Fall time: *NA*
- e) PRF: *0 TO 20 kHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*120/240 V*

*50-60 Hz*

*[Signature]*