


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PAH-5-141-C-1*

S.N.: *4822*

Date: *APRIL 12 89*



40 db ATTEN

10 VOLTS/DIV

PRF ≈ 500 kHz

- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
≤ 350 PSEZ
- c) Rise time: *(FWHM)*
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240V
50-60 Hz

[Signature]