


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MP-AN-1N3-C-M-P-PN*  
S.N.: *4825*  
Date: *JUNE 8 1989*



- a) Output signal amplitude:  
*0 TO ±40 VOLTS*
- b) Pulse width:  
*0.4 TO 2.0 NS*
- c) Rise time:  
*≤ 150 PSEC*
- d) Fall time:  
*≤ 250 PSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*170 / 240 V*  
*50-60 Hz*

*50 db ATTEN*  
*32 VOLTS/DIV*  
*PRF = 100 KHz*

*[Signature]*