


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APP-A-123-C-M-P-PN*

S.N.: *4826*

Date: *JUNE 8 1989*



- a) Output signal amplitude: *0 TO ± 40 VOLTS*
- b) Pulse width: *0.4 TO 2.0 NS*
- c) Rise time: *≤ 150 PSE*
- d) Fall time: *≤ 250 PSE*
- e) PRF: *0 TO 1.0 PMHZ*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz

50 dB ATTEN

32 VOLTS / DIV

PRF = 100 KHZ