


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-2 PS-D-M-P*  
S.N.: *4327*  
Date: *JUNE 8 '89*



- a) Output signal amplitude: *0 TO +15 V.*
- b) Pulse width: *0.2 TO 2.0 NS*
- c) Rise time: *≤ 100 PSE*
- d) Fall time: *≤ 135 PSE*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability:

*40 dB ATTEN*

*10 VOLTS/DIV*

*PRF = 7.0 MHz*

- g) Prime power: *O.K.*  
*120/240 V*  
*50-60 Hz*