


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-5B-PS-P*

S.N.: *4828*

Date: *MAR 30 89*



*10 VOLT/DIV*

*$k_c = 0.7 \Omega$*

*$\approx 14 \text{ AMPS/DIV}$*

*PRF = 700 Hz*

- a) Output signal amplitude:  
*0 TO 140 AMPS TO*
- b) Pulse width: *0.75 ns*  
*25 TO 500 nSEC*
- c) Rise time:  
 *$\leq 10 \text{ nSEC}$*
- d) Fall time:  
 *$\leq 10 \text{ nSEC}$*
- e) PRF:  
*0 TO 1.0 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz*