


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PCR-1-PW-C-P*
S.N.: *4831*
Date: *MMX 3 89*



R_L = 50 Ω
PRF = 1.0 kHz

- a) Output signal amplitude:
0 TO +200 V
- b) Pulse width:
0.1 TO 5.0 μSEC
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
0 TO 1.0 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz

[Signature]