


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-M1-P*

S.N.: *4848-*

Date: *MAR 31 89*



- a) Output signal amplitude:
≥ +30 VOLTS.
- b) Pulse width:
≤ 1 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTEN

10 VOLTS DIV

PRF = 700 KHz

+135 ± 5 VOLTS

5 mA mA