


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-m1-P .

S.N.: 4849

Date: march 31, 1989



- a) Output signal amplitude:
 $\gg +30$ volts
- b) Pulse width:
 ≤ 1 nsec
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
 $+135 \pm 5$ volts
5 mA MAX

40 db atten
 $\therefore 10$ volts/div
PRF = 700 KHz