


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-m1-P .

S.N.: 4850

Date: march 31, 1989



a) Output signal amplitude:

$\geq +30$  volts

b) Pulse width:

$\leq 1$  nsec

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 to 1.0 MHz

f) Jitter, stability:

OK

g) Prime power:

$+135 \pm 5$  volts

5 mA MAX

40 db atten  
 $\therefore 10$  volts / div  
PRF = 700 KHz