


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ACH-M1-N*

S.N.: *4851-*

Date: *MAR 31 1989*



- a) Output signal amplitude:  
*> -30 VOLTS*
- b) Pulse width:  
*≤ 1 NSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK.*
- g) Prime power:

*40 dB ATTEN*

*10 VOLTS DIV*

*PRF = 700 KHz*

*+135 ± 5 VOLTS*

*5 mA (MAX)*

*[Signature]*