


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-m1-N

S.N.: 4852

Date: March 31, 1989



- d) Output signal amplitude:
 ≥ -30 volts
- e) Pulse width:
 ≤ 1 nsec
- f) Rise time:
NA
- g) Fall time:
NA
- h) PRF:
0 to 1.0 MHz
- i) Jitter, stability:
OK
- j) Prime power:
 $+135 \pm 5$ volts
5 mA (MAX)

40 db atten
 $\therefore 10$ volts/div
PRF = 700 KHz