


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-m1-N

S.N.: 4853

Date: march 31, 1989



- a) Output signal amplitude:
 ≥ -30 volts
- b) Pulse width:
 ≤ 1 nsec
- c) Rise time:
 NA
- d) Fall time:
 NA
- e) PRF:
 0 to 1.0 MHz
- f) Jitter, stability:
 OK
- g) Prime power:
 +135 \pm 5 volts
 5 mA (MAX)

40 db attenuation

: 10 volts/div

PRF = 700 KHz

A handwritten signature or scribble in the bottom right corner of the page, consisting of several overlapping loops and lines.