


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PTK-AV-C-P-PN*  
S.N.: *4854*  
Date: *MAR 31 89*



- a) Output signal amplitude:  
*0 TO ± 15 VOLTS*
- b) Pulse width:  
*2 TO 50 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:

*40.66 ATEN*  
*10 VOLTS/DIV*  
*50 KHz 100ns*

*OIL*  
g) Prime power:  
*120 / 240 V*  
*5.0 60 Hz*  
*AA*