


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-1PW-C-N-05-IM*
S.N.: *4855*
Date: *MAY 29 89*



- a) Output signal amplitude:
0 TO - 200 VOLTS
- b) Pulse width:
0.1 TO 5.0 uS
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

$R_L = 50 \Omega$
PRF = 1.0 KHz

120/240 V
50 - 60 Hz