


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-4A-PW-C-2-CA-M1*

S.N.: *4874*

Date: *JULY 21 '89*



R_c = 50 Ω
PRF = 500 Hz

- a) Output signal amplitude:
0 TO +400 V
- b) Pulse width:
0 TO 5.0 μS
- c) Rise time:
≤ 20 nS
- d) Fall time:
≤ 20 nS
- e) PRF:
0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V, 50-60 Hz

$V_{EA} (V_{rms})$	$V_{OUT} (V_{rms})$
1.0	50
2.0	100
3.1	150
4.2	200
5.3	250
6.3	300
7.4	350
8.6	400