


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-C-P*

S.N.: *4875*

Date: *MAY 8 '89*



40 db ATTEN

10 VOLTS/DIV
PRF ≈ 10 MHz

- a) Output signal amplitude: *0 TO +15V*
- b) Pulse width: *0.2 TO 2.0 NSEC*
- c) Rise time: *≤ 100 PSEC*
- d) Fall time: *≤ 100 PSEC*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50 to 1Hz