



PULSE GENERATOR  
PERFORMANCE CHECK

Model: MUR-4B-PW-C-N  
 S.N.: 4877  
 Date: JUNE 19 89



- a) Output signal amplitude:  
0 TO -400V.
- b) Pulse width:  
0.1 TO 100  $\mu$ SEC.
- c) Rise time:  
 $\leq 20$  NSEC
- d) Fall time:  
 $\leq 20$  NSEC

(A)  $R_L = 50\Omega$   
 $PRF \approx 1$  KHz



- e) PRF:  
MARK PW: 0 TO 50 Hz
- f) Jitter, stability:  
PW  $\leq 5$   $\mu$ SEC: 0 TO 1 KHz
- g) Prime power:  
OK  
120/240V  
50-60 Hz

(B)  $R_L = 50\Omega$   
 $PRF \approx 50$  Hz

*[Signature]*