


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMH-3A-N*

S.N.: *4878*

Date: *APRIL 20 1989*



50 db ATTEN

~ 32 VOLTS/DIV

PRF = 7.0 MHz

a) Output signal amplitude:

0 TO -50 VOLTS

b) Pulse width:

≤ 600 PSEC.

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 ± 1.0 VOLTS

600 mA (MAX)