


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-C-P-PN*

S.N.: *4382*

Date: *JUNE 13 89*



a) Output signal amplitude:

*0 TO ± 5 VOLT*

b) Pulse width:

*0.2 TO 6.0 NS*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*120/240 V*

*50-60 Hz.*

*40 db ATTEN  
1.5 VOLT/DIV  
PRF = 12 MHz.*

