


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-V-1A2-C-P*


S.N.: *4838*

Date: *APRIL 27 89*



- a) Output signal amplitude:  
*0 TO +100 VOLTS*
- b) Pulse width:  
*4 TO 100 NSEC*
- c) Rise time:  
*≤ 500 PSEC*
- d) Fall time:  
*≤ 3 NSEC.*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/280V*  
*50-60HZ*

*60 db ATTEN*  
*100 VOLTS/DIV*  
*PRF = 5.0 KHZ*



*60 db ATTEN*  
*100 VOLTS/DIV*  
*PRF = 5.0 KHZ*  
*(RISE TIME)*

*[Signature]*