


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-PS-N-05*

S.N.: *4931*

Date: *11 MAY 11 89*



70 db ATTEN

≈ 320 VOLTS/DIV

PRF ≈ 2.0 KHz

- a) Output signal amplitude:
0 TO -350 V
- b) Pulse width:
5 TO 100 NSER
- c) Rise time:
≈ 2 NSER
- d) Fall time:
≈ 3 NSER
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50-60 Hz