


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-A-1-PW-EW-DA-P*
S.N.: *4934*
Date: *MAY 16 89*



$V_L = 50 \text{ V}$
 $\text{PRF} \approx 5.0 \text{ kHz}$

- a) Output signal amplitude:
0 to +200 V
- b) Pulse width:
50 NSER TO
- c) Rise time: *100 NSER*
≤ 10 NSER
- d) Fall time:
≤ 10 NSER
- e) PRF:
MAX PW: 0 TO 5 kHz
MIN PW: 0 TO 100 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 VOLT
600 mA (MAX)